



Vista 200

IR-PIFM • PIF-IR • AFM-IR

Sub 5 nm IR spatial resolution • Single-molecule-level sensitivity

Vista 200: power meets precision

Controlled environment

A metal enclosure with mylar wrapped insulating foam ensures acoustic isolation and stable temperature control within 0.1 °C, while also being clean room compatible.

Clear signal

Our fully enclosed light path maximizes infrared transmission by eliminating atmospheric interference from water vapor, dramatically improving signal-to-noise ratio (SNR).

Sub-5 nm IR resolution

Exceptional AFM performance

With a 120 µm xy-scanner, 12 µm sample z-scanner, and integrated vibration isolation, our AFM is top notch. The high-bandwidth dual-z feedback system allows true non-contact AFM for IR PiFM.

Ultimate spectro-nanoscscopy

Photo-induced force microscopy (PiFM) and Photo-induced force infrared (PiF-IR) spectroscopy are the leading nano-IR techniques. Non-destructive, non-contact, and with monolayer sensitivity, our instrument is best in class.

Reliable spectra

Active beam steering maintains spectral consistency across the entire sample, and our non-contact AFM prevents cross-contamination.

Dynamic laser control

Our optical multiplexer handles polarization and normalization automatically for effortless laser control, allowing you to choose a light source while keeping everything co-aligned.

Semiconductor-specific

Streamlined processes

Our included data analysis software allows your team to go from measurements to presentation-ready results quickly. Tight integration with our data acquisition software allows you to get the answers you need in real time. Our scripting API allows custom recipes, and our files are open so integration with 3rd party software is easy.

Made for your samples

Choose between a vacuum chuck with magnetic mounts or an EUV mask holder to suit the needs of your lab. This instrument is perfect for 4-, 6-, or 8-in wafers.

Single-molecule-level sensitivity

Powerful yet non destructive

PiFM operates in true non-contact mode. The instrument doesn't touch your pristine samples, so they stay clean – and so does the tip.

Built for industry



200 mm stage

Maximize efficiency minimize disruption

The 300 mm wide sample access door minimizes thermal disturbances that cause drift while allowing you to maximize throughput. With our forward moving stage and user friendly head mount, both the tip and sample can be easily swapped without opening the entire enclosure.

Scientific principles PiFM & PiF-IR




FIGURE 1. A pulsed and tunable IR laser is focused onto the apex of a metal coated AFM tip. The laser is modulated at a frequency carefully calculated based on the resonance frequencies of the cantilever.




FIGURE 2. The metal-coated AFM tip acts as an antenna and creates a highly local enhanced field (yellow). This field locally polarizes the sample, resulting in an attractive force whose magnitude depends on the absorption strength of the sample.




FIGURE 3. One resonance of the cantilever is used to detect the PiFM signal. Simultaneously, another resonance is used to collect the standard AFM topography and phase in a non-contact manner (no tip/sample contamination).




FIGURE 4. The depth of the tip-enhanced illumination depends heavily on the spacing between the tip and the sample. By detecting the attractive forces in non-contact mode, the measurement is made extremely sensitive – capable of detecting monolayers of material.




FIGURE 5. PiF-IR nanoscale spectra are made by measuring the strength of the attractive photo-induced forces as a function of wavenumber. PiFM chemical maps are created by scanning the surface with a fixed-wavenumber to measure absorption strength as a function of position. The color layer over the AFM topography is three fixed-wavenumber PiFM images combined.

PiFM chemical mapping

Reveal hidden structure in color




FIGURE 6. 3D visualization of the AFM topography of a cell wall from an ultra-thin cross section of spruce wood. An overlay of two PiFM images shows the chemical composition of the surface where lignin and cellulose mix. This PiFM overlay reveals how the materials are distributed, and it shows how some of the topographic features are related to the local chemistry. Scan dimensions: $1\text{ }\mu\text{m} \times 1\text{ }\mu\text{m} \times 0.034\text{ }\mu\text{m}$.

Sub-5 nm spatial resolution




FIGURE 7. A zoomed-in region of the spruce wood cell wall. At only 150 nm square, this scan area is over 50 times smaller than in figure 6! PiFM images show the chemical distribution of lignin and cellulose on the surface. A line trace plotting the intensity of the data in the green image shows an IR spatial resolution of less than 5 nm. Notice how the PiFM images bring the topography alive with precise chemical detail while the topography itself is unremarkable. Scan dimensions: $150\text{ nm} \times 150\text{ nm} \times 10.5\text{ nm}$.

PiF-IR nano-spectroscopy

Single-molecule-level sensitivity




FIGURE 8. This sample was intended to be a uniform monolayer of peptoid molecules; however, initial PiF-IR analysis indicated that actual peptoid coverage was very sparse. Therefore, a PiFM image was taken at 1633 cm⁻¹ which should highlight any peptoid molecules present. This image reveals tiny peptoid fragments on the surface. A PiF-IR spectrum taken on one of those points shows the characteristic peak at 1633 cm⁻¹ despite the peptoid fragment being only 1.7 nm tall! The black spot in the green PiFM image shows that the fragment is sitting on top of the substrate material. Scan dimensions: 500 nm × 500 nm × 1.7 nm.

Excellent agreement with FTIR




FIGURE 9. On homogeneous samples, PiF-IR spectra agree with FTIR extremely well.

Analyze organic and inorganic materials




FIGURE 10. Both organic and inorganic samples can be analyzed equally well using PiFM and PiF-IR. Here, a PS-*b*-PMMA block copolymer has undergone sequential infiltration synthesis to produce aluminum oxides in the PMMA block. The sample was exposed to a vapor of trimethyl aluminum, which should only react with carbonyl groups in the PMMA. Subsequent exposure to water vapor converts the trimethyl aluminum into aluminum oxide. PiFM analysis after this process confirms the block-selective infiltration by the presence of a new broad peak from 800 to 1100 cm⁻¹ (not shown). A PiFM image taken at 900 cm⁻¹ (yellow image) highlights the infiltrating alumina in the PMMA blocks, demonstrating the exceptional resolution of PiFM chemical mapping even on inorganic samples. Scan dimensions: 500 nm × 500 nm × 5.5 nm.

Orientation discrimination

Average PiF-IR correlate with FTIR!





FIGURE 11. Similar to polarized FTIR, PiF-IR spectra are orientation sensitive due to the tip-enhanced field. The PiF-IR spectra above were acquired at different crystal faces on a vitamin C sample. They exhibit the local IR bands associated with the specific crystal faces. The average of these three distinct local PiF-IR spectra compare nicely with the bulk FTIR spectrum, which is the equivalent of an ensemble average of billions of PiF-IR spectra. Scan dimensions: 640 nm × 640 nm × 1025 nm.

Comparisons

Comparing surface analytical techniques

	PiFM & PiF-IR	Raman	FTIR	TOF-SIMS	XPS	TXRF	SEM/EDS	TEM	Auger
Species Detected	Molecular	Molecular	Molecular	Molecular	Molecular	Elemental	Elemental	Elemental	Elemental
Chemical Mapping	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes
Lateral Resolution	Sub 5 nm	> 0.5 μm	> 10 μm	100 nm	Sub 5 μm	~10 nm	1 nm *0.5 μm EDS	0.2 nm *1 – 20 nm EDS	8 nm
Depth Probed	20 nm & bulk	> 500 nm	1 μm	1 nm	10 nm	10 nm	1 μm	~100 nm	10 nm

TABLE 1. PiFM & PiF-IR bring molecular analysis to the realm of true nanoscale resolution, providing both IR absorption spectra and chemical mapping with sub-5 nm spatial resolution and monolayer sensitivity, complementing XPS or ToF-SIMS in any high-end lab. *For SEM and TEM, EDS measurements are not as high resolution as is possible for imaging.

Vista 200 specifications

Stage and scanner

Sample stage travel	200 mm × 200 mm square.
Scan size	120 × 120 μm .
Dual Z feedback	12 μm z-scanner with 600 nm fast z-scanner provides both high bandwidth and a large z-range.

Physical requirements

Table size	1.2 m × 1.8 m (4 ft × 6 ft) optical breadboard recommended.
Enclosure	300 mm × 200 mm access door minimizes thermal drift. The enclosure is removable without disconnecting cables.

Functionality

Imaging modes	Non-contact AFM, PiFM, KPFM, cAFM, nano DMA, FvD (force vs distance) mapping.
Spectroscopy modes	PiF-IR, FvD.
PiF Laser options	QCL (770 – 1840, 1995 – 2395 cm^{-1}), OPO/DFG (590 – 2050, 2250 – 4400, 5000 – 7000 cm^{-1}).
Depth Probed	20 nm in surface mode & greater than 100 nm in bulk mode.